

Patient name: Sample Patient
DOB: 01/22/1965
Follow-up type: 30 day follow-up
Person interviewed: Sample


Date: 03/04/2017
Doctor: Marcus Welby, M.D.
Setup date: 02/04/2017

Compliance Information

Date range	02/04/2017 - 03/04/2017 (29 days)
Percent days with device usage	93.1% (27 days)
Average usage (all days)	5 hours, 54 minutes, 36 seconds
Percent of days with usage >= 4 hours	65.5%

Sleep Therapy Statistics (Philips Respironics)

Average time in large leak per day	2 minutes, 24 seconds
Average AHI	4.1
CPAP pressure	6-20



Digitally signed by: Sample Employee



Patient name: Sample Patient	Date: 03/04/2017
DOB: 01/22/1965	Doctor: Marcus Welby, M.D.
Follow-up type: 30 day follow-up	Setup date: 02/04/2017
Person interviewed: Sample	

Current pressure setting	6-20cm
Current mask	Swift LT-Medium
Patient's average sleep time with device?	6-9 hours
Does the patient use the machine nightly?	No
If no, how often is device used?	Missed four days due to mask comfort issues
How does the patient compare their level of fatigue since the start of therapy?	Less fatigue
Does the patient have any of the following problems?	
Difficulty falling asleep or staying asleep?	No
Uncomfortable fit of mask, nasal pillows, seals, or headgear?	Yes
Irritated skin, pressure sores, etc.?	No
Runny nose/congestion, irritated eyes, or dry throat?	No
Taking mask off while sleeping?	Yes
Is patient still snoring since the start of use of the device?	No
Service evaluation	
Does the patient know how to contact Classic SleepCare for service?	Yes
Does the patient require any supplies at this time?	Yes
Does the patient know the recommended schedule for changing disposable supplies?	Yes
Does the patient know the recommended cleaning procedure?	Yes
Does the patient understand the instructions for safety and use of the device?	Yes

Call notes

Patient is using her therapy most nights. She does state that she is having problems with the fit of her mask, which is why on some nights she is not able to sleep 4 hours with the machine. In reviewing proper fit and placement, patient revealed that she had not been cleaning her nasal pillow daily. Discussed that this may be the issue with the comfort of her mask on some nights. Patient states that she will start cleaning daily and realizes that this has most likely been the issue. CSC will follow up again at her 90-day mark. Patient knows to call if cleaning recommendation does not solve the issue. If this occurs, we will change to another style of mask.

